



Łukasiewicz
GIT

Łukasiewicz Research Network
– Upper Silesian Institute of Technology

44-100 Gliwice, ul. Karola Miarki 12-14, POLAND
Tel.: +48 32 23 45 251
E-mail: rm@git.lukasiewicz.gov.pl



POLCARGO
MEDYKA
POLCARGO – MEDYKA Sp. z o.o.
International Superintendence
& Testing Services, Poland

CERTIFICATE OF ANALYSIS

PI 3.21

REFERENCE MATERIAL OF IRON ORE CONCENTRATE J.GOK

Analysis listed as percent by weight [% m/m]

Fe	64.94	K₂O	0.029
FeO	25.94	Na₂O	0.077
SiO₂	8.33	V	0.0005
CaO	0.15	Cr	(0.002)
Mn	0.017	Co	0.0009
Al₂O₃	0.20	Ni	0.0024
TiO₂	0.016	Zn	(0.003)
MgO	0.44	Cl	0.083
C	0.18	Ba	0.0019
P	0.015	GOI	1.99
S	0.026		

Values in brackets are informative

Certificate Number: PI 3.21-081124
Certificate revision history on page 4

Analysis	Fe	FeO	SiO ₂	CaO	Mn	Al ₂ O ₃	TiO ₂	MgO
1	64.887	25.697	8.124	0.1560	0.0150	0.179	0.013	0.405
2	64.898	25.757	8.134	0.1370	0.0167	0.187	0.014	0.426
3	64.900	25.924	8.256	0.1470	0.0169	0.193	0.015	0.449
4	64.903	25.948	8.293	0.1620	0.0170	0.198	0.016	0.452
5	64.910	26.073	8.390	0.1580	0.0183	0.209	0.016	0.453
6	64.957	26.217	8.460	0.1645	0.0202	0.220	0.017	0.465
7	65.003		8.477				0.018	
8	65.005		8.490					
9	65.012							
Average	64.942	25.936	8.328	0.1541	0.0174	0.198	0.016	0.442
Std. Dev.	0.053	0.193	0.149	0.0103	0.0017	0.01	0.002	0.022
Certified	64.94	25.94	8.33	0.15	0.017	0.20	0.016	0.44
C(95%)	0.04	0.22	0.13	0.01	0.002	0.017	0.002	0.025

Analysis	P	S	K ₂ O	Na ₂ O	Pb	Cu	Ni	Ba
1	0.0127	0.0253	0.0261	0.0683	0.0011	0.0002	0.0020	0.0016
2	0.0131	0.0257	0.0267	0.0720	0.0024	0.0016	0.0021	0.0017
3	0.0134	0.0264	0.0270	0.0770			0.0026	0.0020
4	0.0148	0.0265	0.0275	0.0803			0.0027	0.0022
5	0.0174	0.0267	0.0327	0.0820				
6	0.0174	0.0277	0.0327	0.0847				
7	0.0182		0.0333					
Average	0.0153	0.0264	0.0294	0.0774			0.0024	0.0019
Std. Dev.	0.0023	0.0008	0.0033	0.0062			0.0004	0.0003
Certified	0.015	0.026	0.029	0.077			0.0024	0.0019
C(95%)	0.0023	0.0010	0.0033	0.0072			0.0006	0.0005

Analysis	Cr	Cl	C	Zn	Co	V	GOI
1	0.0023	0.0795	0.1747	0.0017	0.0006	0.0003	1.920
2	0.0023	0.0810	0.1760	0.0022	0.0007	0.0004	1.950
3	0.0025	0.0830	0.1790	0.0029	0.0009	0.0004	1.980
4		0.0870	0.1850	0.0037	0.0010	0.0006	2.020
5			0.1868		0.0012	0.0008	2.023
6			0.1900				2.030
Average	0.0024	0.0826	0.182	0.0026	0.0009	0.0005	1.987
Std. Dev.	0.0001	0.0032	0.0062	0.0009	0.0002	0.0002	0.0450
Certified		0.083	0.18		0.0009	0.0005	1.99
C(95%)		0.006	0.009		0.0003	0.0003	0.05

$C(95\%) = (t \cdot sd) / \sqrt{n-1}$ - The half-width confidence interval, calculated for the 95 % confidence level, where t is the appropriate Student's t value, sd is the interlaboratory standard deviation and n is the number of acceptable mean values. For further information regarding the confidence interval for the certified value see ISO Guide 35:1989 section 4.

Certification Process: Both preparation of this Reference Material and certification process were prepared according to requirements of ISO Guide 31, ISO Guide 34 and ISO Guide 35. This Reference Material is in agreement with ISO Guide 30.

Chemical Analysis: Chemical analyses were carried out on dried at 105°C powder samples. Single values in the above table are the means obtained by individual laboratories. The following methods were used for analysis:

Fe - titrimetric, XRF;
FeO - titrimetric;
SiO₂ - gravimetric, photometric as silicon-molybdenum blue, XRF;
CaO - titrimetric, ICP-AES, XRF;
Mn - flame AAS, ICP-AES, XRF;
Al₂O₃ - flame AAS, ICP-AES, XRF;
TiO₂ - flame AAS, ICP-AES, XRF;
MgO - ICP-AES, XRF;
C,S - high frequency infra-red absorption (HFIR), XRF;
P - ICP-AES, photometric as molybdenum blue, XRF;
K₂O - AAS, ICP-AES, XRF;
Na₂O - AAS, ICP-AES, XRF;
GOI - gravimetric;
Cl - spectrophotometric,
V, Cr, Co, Ni, Zn, Pb, Cu, Ba - ICP-AES, GF-AAS.

The laboratories participating in the testing of this Reference Material:

Mittal Steel Poland S.A. Kraków, Poland
Mittal Steel Poland S.A. Dąbrowa Górnicza, Poland
Mittal Steel Ostrava, Czech Republic
Huta Trzyniec, Czech Republic
U.S. Steel, Koszyce, Slovakia
Polcarga – Medyka, Poland
Instytut Metalurgii Żelaza, Poland

Homogeneity: The homogeneity of this Reference Material was evaluated with the use of X-ray fluorescence spectrometry and found acceptable.

Traceability: This Reference Material was tested with the use of UV-Vis spectrometry, AAS, ICP-AES and C,S-analyzers and was found compatible to the following CRMs: ECRM 676-1 , ECRM 683-1, ASCRM-004 ÷ 0007, JSS 800 ÷ 805.

Origin of the material: Ukraine

Basic mineralogical composition:

Main component: magnetite Fe₃O₄

Other components: hematite α - Fe₂O₃, quartz SiO₂, siderite (Fe,Mg)CO₃, magnesite MgCO₃, dolomite CaMg(CO₃)₂, chlorite

Available form: 100 g of powder sample, grain size less than 0.1 mm.

Intended use: This Reference Material is intended for use in determination of chemical composition of iron ores by x-ray fluorescence spectrometry, UV-Vis spectrometry, AAS, ICP-AES and C,S-analyzers and other wet methods. Chemical analyses should be carried out on samples dried at 105°C.

Validity of certification: The certification of PI 3.21 is valid for 15 years - until February 2035, within the uncertainty specified provided this Reference Material is stored in accordance with the instructions given in this certificate (see Storage). The certification is nullified if this Reference Material is damaged, contaminated or otherwise modified.

Revision: This Reference Material was certified originally in February 2005. Additional tests were performed to prove that the material remains unchanged.

Storage: This Reference Material should be stored in dry place and in environment free from chemical or other aggressive vapours and should be protected against vibration. If the contents become changed (for example oxidized) because of contamination, the whole contents of bottle should be discarded.

Safety: This Reference Material and packing do not contain substances which can directly influence health. Radioactivity less than 0.12 Bq/g equivalent of ^{60}Co

Inquiries regarding this Reference Material should be directed to rm@git.lukasiewicz.gov.pl


Approved by

President of Polcargo Medyka



Marek Drabik, BEng, MSc

Director of the Institute



Prof. Dr. Hab. Eng. Adam Zieliński

Certificate issue date: 08.11.2024

Certificate revision history:

08 November 2024 (editorial changes)

25 February 2020 (change of information regarding validity of certification, editorial changes);

February 2005 (Original certificate date)